Series Fetal & Maternal Monitor

Support Solutions




Specification

Physical Characteristics						Display Specifications					
Dimensions: 347mm x 330mm x 126mm Weight: Approx. 6 kg									2.1 inches color TFT-LCD RGB 800 (W)×600 (H)		
Ultrasound						DECG					
Technique: Ultrasound Pulse Doppler with autocorrelation Pulse Repetition Rate: 2 KHz Ultrasound Frequency: 1.0 MHz ±10% FHR Measurement Range: 50 bpm ~ 240 bpm Resolution: 1 bpm Accuracy: ±1 bpm /ob: <10 mW/cm²						Technique: Peak-peak detection technique DFHR Measurement Range: 30 bpm ~ 240 bpm Resolution: 1 bpm Accuracy: ±1 bpm					
TOCO						IUP					
OCO Range: 0% ~ 100% Non-linear Error: ≤10% Resolution: 1% Zero Mode: Automatic / Manual						Pressure Range: 0 ~ 100 mmHg Non-linear Error: ≤± 3 mmHg Resolution: 1% Zero Mode: Automatic / Manual					
AFM	Duland Day					ECG					
Technique: Range: Resolution:	Pulsed Doppler ultrasound 0 ~ 100 (%) 1%					Manual control ECG waveform display ECG falls off: Detect automatically					
HR						SpO ₂					
easurement Range: 30 bpm ~ 240 bpm easuring Accuracy: ±2 bpm R					Measurement Range: 50% ~ 100% Resolution: 1% NIBP (for adult)						
Measurement Range: Measuring Accuracy:	30 bpm ~ 240 bpm ±2 bpm					Systolic pressure: Mean pressure: Diastolic pressure: Resolution:		40 mm 20 mm 10 mm	40 mmHg ~ 270 mmHg 20 mmHg ~ 235 mmHg 10 mmHg ~ 215 mmHg 1 mmHg		
TEMP						Recorde	er				
Measurement Range: 0 °C ~ 50 °C Accuracy: ±0.2 °C						Recording Paper: Z-fold, 150/152mm thermosensitive paper Recording Speed: 1/2/3 cm/min, 25 mm/sec for history data					
Power Supply											
Operating Voltage: Operating Frequency: Battery:	100~240V~ 50/60 Hz 14.8V/4400		echarge	eable Li-ion ba	ttery						
Configuration Twins	FHR TOCO	FM	AFM	DECG/IUP	MECG	NIBP	MSPO ₂	TEMP	Fetal Stimulator	Lithium-ion Batte	
F9 🗸	· /	V	V	Opt	×	×	×	×	Opt	Opt	
EQ Eypross	V	./		Ont	1	1/	1/	./	Opt	Opt	

EDAN Edan Instruments, Inc.

3/F - B, Nanshan Medical Equipment Park, Nanhai Rd 1019#, Shekou, Nanshan Shenzhen, 518067 P.R. China Tel: +68-755-26898326 Fax: +86-755-26898330 www.edan.com.cn Email: info@edan.com.cn





Wall Mount

EDAN





EDAN

Fetal & Maternal Monitor

Quick setup and simplified work flow

Functional keys and knob provide various shortcuts to achieve functions for clinical use. The 'start' button could be configured to integrate patient info and printing. This could help the doctor to simplify the workflow and work with 1 button only.

















Various printing options

F9 series are compatible with 152 and 150 mm paper in either American or International standard. The recording information includes patient info, FHR waveform, TOCO waveform, Fetal movement, maternal heart rate and SpO2 traces and so on.

Telemetry transducer system

Wireless monitoring offers full range of comfort and convenience, the lightweight and waterproof transducers gives the patient maximized flexibility as well as convenience for clinician. As an optional accessory of F9, the telemetry transducer system perfectly suits a wide range of OB monitoring applications, including both antepartum and intrapartum.

F9 series fetal & maternal monitor provide solutions for all areas of fetal monitoring, ranging from antepartum, intrapartum, and postpartum clinical applications. With integrated monitoring of twins' FHR, uterine activity, fetal movement, intrauterine pressure and direct ECG as well as maternal NIBP, SpO₂, ECG and TEMP, F9 series offer the most cost effective and flexible solution for all your fetal monitoring needs, without compromising quality, precision, performance and ergonomics.

Multi-display Modes



Switching among three display modes, user may choose the most suitable one for clinical use that allows care givers to evaluate maternal data simultaneously

The touch-screen, color display allows quick setup and optimized performance.


Max 60°

12.1 inches color touch screen





12- crystal and 1MHz waterproof transducer



Our 12-crystal ultrasound probe employs advanced technology unique to EDAN, providing wider beam area and more homogeneous signal, thus realizing better performance for bedside monitors than conventional transducers.